

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022725\
 Data File : VY021351.D
 Acq On : 27 Feb 2025 12:28
 Operator : SY/MD
 Sample : Q1442-03
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 351

Quant Time: Feb 28 00:41:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 26 02:09:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	118963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	213600	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	193897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	79357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	76242	56.495	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	113.000%	
35) Dibromofluoromethane	7.634	113	74099	52.914	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	105.820%	
50) Toluene-d8	10.103	98	257396	47.559	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	95.120%	
62) 4-Bromofluorobenzene	12.408	95	79764	43.793	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	87.580%	
Target Compounds						Qvalue
20) Methylene Chloride	4.610	84	10933	7.810	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022725\
Data File : VY021351.D
Acq On : 27 Feb 2025 12:28
Operator : SY/MD
Sample : Q1442-03
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
351

Quant Time: Feb 28 00:41:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022525S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 26 02:09:13 2025
Response via : Initial Calibration

