

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022824\
 Data File : VY017512.D
 Acq On : 28 Feb 2024 11:11
 Operator : SY/MD
 Sample : VY0228SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0228SBL01

Quant Time: Feb 28 23:10:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	135700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	236723	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	204484	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	76214	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	62530	48.621	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.240%
35) Dibromofluoromethane	7.722	113	74464	45.142	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.280%
50) Toluene-d8	10.185	98	267534	46.387	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.780%
62) 4-Bromofluorobenzene	12.483	95	69319	38.244	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	76.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022824\
Data File : VY017512.D
Acq On : 28 Feb 2024 11:11
Operator : SY/MD
Sample : VY0228SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0228SBL01

Quant Time: Feb 28 23:10:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

