

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022824\
 Data File : VY017518.D
 Acq On : 28 Feb 2024 14:01
 Operator : SY/MD
 Sample : P1616-03
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 146-BAY-AVENUE

Quant Time: Feb 28 23:13:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	2332	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	3189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	2337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	263	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1124	50.857	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.720%
35) Dibromofluoromethane	7.728	113	817	36.765	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	73.540%
50) Toluene-d8	10.185	98	3352	43.142	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	86.280%
62) 4-Bromofluorobenzene	12.483	95	429	17.569	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	35.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022824\
Data File : VY017518.D
Acq On : 28 Feb 2024 14:01
Operator : SY/MD
Sample : P1616-03
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
146-BAY-AVENUE

Quant Time: Feb 28 23:13:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

