

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022824\
 Data File : VY017520.D
 Acq On : 28 Feb 2024 14:49
 Operator : SY/MD
 Sample : P1616-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 196-BAY-AVENUE

Quant Time: Feb 28 23:13:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	109581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	198981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	180137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	67648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	53894	51.894	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.780%
35) Dibromofluoromethane	7.709	113	65965	47.575	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.140%
50) Toluene-d8	10.185	98	227456	46.918	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.840%
62) 4-Bromofluorobenzene	12.483	95	59700	39.184	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.360%
Target Compounds						
17) Carbon Disulfide	4.186	76	3796	1.222	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022824\
Data File : VY017520.D
Acq On : 28 Feb 2024 14:49
Operator : SY/MD
Sample : P1616-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
196-BAY-AVENUE

Quant Time: Feb 28 23:13:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

