



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022824\
Data File : VY017510.D
Acq On : 28 Feb 2024 08:26
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Title : SW846 8260
Last Update : Sat Feb 24 04:41:20 2024



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	8023	PASS
75	95	30	60	48.5	23920	PASS
95	95	100	100	100.0	49280	PASS
96	95	5	9	6.6	3270	PASS
173	174	0.00	2	1.3	604	PASS
174	95	50	100	97.2	47912	PASS
175	174	5	9	7.2	3457	PASS
176	174	95	101	97.9	46920	PASS
177	176	5	9	6.5	3053	PASS