

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
 Data File : VY017528.D
 Acq On : 29 Feb 2024 15:11
 Operator : SY/MD
 Sample : VY0229SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0229SBL01

Quant Time: Mar 01 00:29:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	118917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	198058	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	167899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	70569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	57390	50.922	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.840%
35) Dibromofluoromethane	7.722	113	69023	50.012	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.020%
50) Toluene-d8	10.185	98	249573	51.720	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.440%
62) 4-Bromofluorobenzene	12.489	95	66218	43.665	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.320%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.716	84	17666	9.296	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
Data File : VY017528.D
Acq On : 29 Feb 2024 15:11
Operator : SY/MD
Sample : VY0229SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0229SBL01

Quant Time: Mar 01 00:29:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

