

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
 Data File : VY017531.D
 Acq On : 29 Feb 2024 16:30
 Operator : SY/MD
 Sample : P1622-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-022824

Quant Time: Mar 01 00:32:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	125456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	224241	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	197587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	57282	48.177	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.360%
35) Dibromofluoromethane	7.722	113	72015	46.087	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.180%
50) Toluene-d8	10.185	98	257908	47.207	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.420%
62) 4-Bromofluorobenzene	12.483	95	67213	39.146	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.300%
Target Compounds						
20) Methylene Chloride	4.723	84	20957	11.219	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
 Data File : VY017531.D
 Acq On : 29 Feb 2024 16:30
 Operator : SY/MD
 Sample : P1622-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-022824

Quant Time: Mar 01 00:32:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

