

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
 Data File : VY017535.D
 Acq On : 29 Feb 2024 18:04
 Operator : SY/MD
 Sample : P1659-07
 Misc : 4.85g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-7-VOC

Quant Time: Mar 01 00:33:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	123416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	223840	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	200339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	75285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	62973	53.839	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.680%
35) Dibromofluoromethane	7.728	113	72305	46.356	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.720%
50) Toluene-d8	10.185	98	257964	47.302	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.600%
62) 4-Bromofluorobenzene	12.489	95	68731	40.102	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.200%
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	21696	12.128	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
Data File : VY017535.D
Acq On : 29 Feb 2024 18:04
Operator : SY/MD
Sample : P1659-07
Misc : 4.85g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-7-VOC

Quant Time: Mar 01 00:33:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

