

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
 Data File : VY017538.D
 Acq On : 29 Feb 2024 19:16
 Operator : SY/MD
 Sample : P1620-31
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-4

Quant Time: Mar 01 00:34:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	34724	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	47478	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.502	117	33921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	9344	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	10965	33.319	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	66.640%
35) Dibromofluoromethane	7.728	113	14730	44.523	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.040%
50) Toluene-d8	10.191	98	50917	44.017	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.040%
62) 4-Bromofluorobenzene	12.495	95	9160	25.197	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	50.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY022924\
Data File : VY017538.D
Acq On : 29 Feb 2024 19:16
Operator : SY/MD
Sample : P1620-31
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-4

Quant Time: Mar 01 00:34:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

