

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012747.D
 Acq On : 01 Mar 2023 13:40
 Operator : KP/MD
 Sample : 01727-01
 Misc : 4.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JB-1S

Quant Time: Mar 02 02:04:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	110600	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	182547	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	159703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	75578	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	72455	55.469	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.940%
35) Dibromofluoromethane	7.716	113	72385	58.707	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.420%
50) Toluene-d8	10.179	98	248684	53.836	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.680%
62) 4-Bromofluorobenzene	12.483	95	79527	53.800	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
Data File : VY012747.D
Acq On : 01 Mar 2023 13:40
Operator : KP/MD
Sample : 01727-01
Misc : 4.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JB-1S

Quant Time: Mar 02 02:04:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

