

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012749.D
 Acq On : 01 Mar 2023 14:54
 Operator : KP/MD
 Sample : 01725-01RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-022723RE

Quant Time: Mar 02 02:05:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	49419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	78825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	71142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	31757	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	37779	64.728	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.460%
35) Dibromofluoromethane	7.716	113	37005	69.505	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	139.000%
50) Toluene-d8	10.179	98	122288	61.309	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	122.620%
62) 4-Bromofluorobenzene	12.483	95	37761	59.159	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
Data File : VY012749.D
Acq On : 01 Mar 2023 14:54
Operator : KP/MD
Sample : 01725-01RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-022723RE

Quant Time: Mar 02 02:05:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

