

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012741.D
 Acq On : 01 Mar 2023 10:31
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 02:01:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.514	0.463	9.9	86	0.00
3 P	Chloromethane	0.687	0.538	21.7	77	0.00
4 C	Vinyl Chloride	0.568	0.516	9.2#	84	0.00
5 T	Bromomethane	0.311	0.306	1.6	93	0.00
6 T	Chloroethane	0.362	0.320	11.6	87	0.00
7 T	Trichlorofluoromethane	0.916	0.880	3.9	90	0.00
8 T	Diethyl Ether	0.328	0.315	4.0	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.547	0.558	-2.0	95	0.00
10 T	Methyl Iodide	0.716	0.723	-1.0	91	0.00
11 T	Tert butyl alcohol	0.065	0.052	20.0	80	0.00
12 CM	1,1-Dichloroethene	0.528	0.518	1.9#	91	0.00
13 T	Acrolein	0.082	0.076	7.3	75	0.00
14 T	Allyl chloride	1.025	0.913	10.9	80	0.00
15 T	Acrylonitrile	0.174	0.170	2.3	81	0.00
16 T	Acetone	0.193	0.225	-16.6	96	0.00
17 T	Carbon Disulfide	1.778	1.400	21.3	73	0.00
18 T	Methyl Acetate	0.621	0.555	10.6	75	0.00
19 T	Methyl tert-butyl Ether	1.421	1.417	0.3	86	0.00
20 T	Methylene Chloride	0.734	0.606	17.4	88	0.00
21 T	trans-1,2-Dichloroethene	0.575	0.565	1.7	90	0.00
22 T	Diisopropyl ether	1.913	1.867	2.4	86	0.00
23 T	Vinyl Acetate	1.185	1.133	4.4	81	0.00
24 P	1,1-Dichloroethane	1.061	1.037	2.3	89	0.00
25 T	2-Butanone	0.256	0.260	-1.6	83	0.00
26 T	2,2-Dichloropropane	0.925	0.903	2.4	93	0.00
27 T	cis-1,2-Dichloroethene	0.629	0.643	-2.2	94	0.00
28 T	Bromochloromethane	0.425	0.412	3.1	83	0.00
29 T	Tetrahydrofuran	0.156	0.146	6.4	75	0.00
30 C	Chloroform	1.028	1.058	-2.9#	94	0.00
31 T	Cyclohexane	1.026	0.845	17.6	80	0.00
32 T	1,1,1-Trichloroethane	0.930	0.924	0.6	92	0.00
33 S	1,2-Dichloroethane-d4	0.591	0.527	10.8	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.338	0.341	-0.9	88	0.00
36 T	1,1-Dichloropropene	0.535	0.514	3.9	88	0.00
37 T	Ethyl Acetate	0.361	0.313	13.3	70	0.00
38 T	Carbon Tetrachloride	0.554	0.564	-1.8	93	0.00
39 T	Methylcyclohexane	0.618	0.598	3.2	86	0.00
40 TM	Benzene	1.494	1.530	-2.4	92	0.00
41 T	Methacrylonitrile	0.176	0.166	5.7	87	0.00
42 TM	1,2-Dichloroethane	0.463	0.452	2.4	86	0.00
43 T	Isopropyl Acetate	0.624	0.587	5.9	77	0.00
44 TM	Trichloroethene	0.386	0.403	-4.4	95	0.00
45 C	1,2-Dichloropropane	0.384	0.398	-3.6#	92	0.00
46 T	Dibromomethane	0.212	0.225	-6.1	91	0.00
47 T	Bromodichloromethane	0.511	0.535	-4.7	94	0.00
48 T	Methyl methacrylate	0.291	0.280	3.8	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012741.D
 Acq On : 01 Mar 2023 10:31
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 02:01:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	79	0.00
50 S	Toluene-d8	1.265	1.249	1.3	87	0.00
51 T	4-Methyl-2-Pentanone	0.337	0.331	1.8	78	0.00
52 CM	Toluene	0.885	0.952	-7.6#	95	0.00
53 T	t-1,3-Dichloropropene	0.534	0.567	-6.2	93	0.00
54 T	cis-1,3-Dichloropropene	0.604	0.637	-5.5	92	0.00
55 T	1,1,2-Trichloroethane	0.288	0.315	-9.4	97	0.00
56 T	Ethyl methacrylate	0.394	0.430	-9.1	89	0.00
57 T	1,3-Dichloropropane	0.507	0.536	-5.7	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.199	-1.5	79	0.00
59 T	2-Hexanone	0.239	0.253	-5.9	82	0.00
60 T	Dibromochloromethane	0.349	0.386	-10.6	97	0.00
61 T	1,2-Dibromoethane	0.276	0.299	-8.3	94	0.00
62 S	4-Bromofluorobenzene	0.405	0.423	-4.4	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.364	0.377	-3.6	98	0.00
65 PM	Chlorobenzene	1.072	1.137	-6.1	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.434	-9.3	101	0.00
67 C	Ethyl Benzene	1.917	2.040	-6.4#	99	0.00
68 T	m/p-Xylenes	0.731	0.780	-6.7	98	0.00
69 T	o-Xylene	0.677	0.743	-9.7	99	0.00
70 T	Styrene	1.135	1.270	-11.9	101	0.00
71 P	Bromoform	0.254	0.276	-8.7	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.695	3.871	-4.8	102	0.00
74 T	N-amyl acetate	1.215	1.148	5.5	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.831	0.842	-1.3	95	0.00
76 T	1,2,3-Trichloropropane	0.623	0.559	10.3	85	0.00
77 T	Bromobenzene	0.885	0.907	-2.5	100	0.00
78 T	n-propylbenzene	4.622	4.778	-3.4	101	0.00
79 T	2-Chlorotoluene	2.568	2.634	-2.6	101	0.00
80 T	1,3,5-Trimethylbenzene	3.086	3.190	-3.4	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.294	0.305	-3.7	95	0.00
82 T	4-Chlorotoluene	2.663	2.716	-2.0	101	0.00
83 T	tert-Butylbenzene	2.636	2.771	-5.1	101	0.00
84 T	1,2,4-Trimethylbenzene	3.017	3.131	-3.8	100	0.00
85 T	sec-Butylbenzene	4.010	4.231	-5.5	104	0.00
86 T	p-Isopropyltoluene	3.303	3.495	-5.8	103	0.00
87 T	1,3-Dichlorobenzene	1.734	1.820	-5.0	104	0.00
88 T	1,4-Dichlorobenzene	1.758	1.783	-1.4	102	0.00
89 T	n-Butylbenzene	3.163	3.328	-5.2	103	0.00
90 T	Hexachloroethane	0.670	0.707	-5.5	107	0.00
91 T	1,2-Dichlorobenzene	1.549	1.620	-4.6	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.137	0.133	2.9	90	0.00
93 T	1,2,4-Trichlorobenzene	0.979	1.012	-3.4	102	0.00
94 T	Hexachlorobutadiene	0.591	0.610	-3.2	106	0.00
95 T	Naphthalene	1.979	2.074	-4.8	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
Data File : VY012741.D
Acq On : 01 Mar 2023 10:31
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 02 02:01:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.867	0.915	-5.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6