

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012750.D
 Acq On : 01 Mar 2023 16:56
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 02:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.514	0.445	13.4	80	0.00
3 P	Chloromethane	0.687	0.499	27.4#	69	0.00
4 C	Vinyl Chloride	0.568	0.482	15.1#	76	0.00
5 T	Bromomethane	0.311	0.283	9.0	84	0.00
6 T	Chloroethane	0.362	0.307	15.2	81	0.00
7 T	Trichlorofluoromethane	0.916	0.859	6.2	84	0.00
8 T	Diethyl Ether	0.328	0.286	12.8	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.547	0.539	1.5	89	0.00
10 T	Methyl Iodide	0.716	0.664	7.3	80	0.00
11 T	Tert butyl alcohol	0.065	0.048	26.2#	70	0.00
12 CM	1,1-Dichloroethene	0.528	0.491	7.0#	83	0.00
13 T	Acrolein	0.082	0.059	28.0#	57	0.00
14 T	Allyl chloride	1.025	0.948	7.5	80	0.00
15 T	Acrylonitrile	0.174	0.155	10.9	71	0.00
16 T	Acetone	0.193	0.181	6.2	75	0.00
17 T	Carbon Disulfide	1.778	1.360	23.5	69	0.00
18 T	Methyl Acetate	0.621	0.507	18.4	67	0.00
19 T	Methyl tert-butyl Ether	1.421	1.344	5.4	79	0.00
20 T	Methylene Chloride	0.734	0.597	18.7	84	0.00
21 T	trans-1,2-Dichloroethene	0.575	0.550	4.3	84	0.00
22 T	Diisopropyl ether	1.913	1.924	-0.6	86	0.00
23 T	Vinyl Acetate	1.185	1.114	6.0	77	0.00
24 P	1,1-Dichloroethane	1.061	1.064	-0.3	88	0.00
25 T	2-Butanone	0.256	0.216	15.6	67	0.00
26 T	2,2-Dichloropropane	0.925	0.900	2.7	90	0.00
27 T	cis-1,2-Dichloroethene	0.629	0.621	1.3	88	0.00
28 T	Bromochloromethane	0.425	0.408	4.0	79	0.00
29 T	Tetrahydrofuran	0.156	0.128	17.9	63	0.00
30 C	Chloroform	1.028	1.017	1.1#	87	0.00
31 T	Cyclohexane	1.026	0.859	16.3	79	0.00
32 T	1,1,1-Trichloroethane	0.930	0.907	2.5	88	0.00
33 S	1,2-Dichloroethane-d4	0.591	0.509	13.9	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.338	0.340	-0.6	85	0.00
36 T	1,1-Dichloropropene	0.535	0.502	6.2	84	0.00
37 T	Ethyl Acetate	0.361	0.276	23.5	60	0.00
38 T	Carbon Tetrachloride	0.554	0.544	1.8	87	0.00
39 T	Methylcyclohexane	0.618	0.592	4.2	83	0.00
40 TM	Benzene	1.494	1.482	0.8	86	0.00
41 T	Methacrylonitrile	0.176	0.141	19.9	72	0.00
42 TM	1,2-Dichloroethane	0.463	0.431	6.9	80	0.00
43 T	Isopropyl Acetate	0.624	0.531	14.9	68	0.00
44 TM	Trichloroethene	0.386	0.388	-0.5	89	0.00
45 C	1,2-Dichloropropane	0.384	0.382	0.5#	85	0.00
46 T	Dibromomethane	0.212	0.204	3.8	80	0.00
47 T	Bromodichloromethane	0.511	0.510	0.2	87	0.00
48 T	Methyl methacrylate	0.291	0.252	13.4	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012750.D
 Acq On : 01 Mar 2023 16:56
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 02 02:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	71	0.00
50 S	Toluene-d8	1.265	1.243	1.7	84	0.00
51 T	4-Methyl-2-Pentanone	0.337	0.290	13.9	66	0.00
52 CM	Toluene	0.885	0.920	-4.0#	89	0.00
53 T	t-1,3-Dichloropropene	0.534	0.530	0.7	84	0.00
54 T	cis-1,3-Dichloropropene	0.604	0.607	-0.5	85	0.00
55 T	1,1,2-Trichloroethane	0.288	0.295	-2.4	88	0.00
56 T	Ethyl methacrylate	0.394	0.391	0.8	78	0.00
57 T	1,3-Dichloropropane	0.507	0.497	2.0	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.188	4.1	73	0.00
59 T	2-Hexanone	0.239	0.211	11.7	66	0.00
60 T	Dibromochloromethane	0.349	0.360	-3.2	88	0.00
61 T	1,2-Dibromoethane	0.276	0.273	1.1	83	0.00
62 S	4-Bromofluorobenzene	0.405	0.419	-3.5	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.364	0.359	1.4	90	0.00
65 PM	Chlorobenzene	1.072	1.090	-1.7	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.410	-3.3	92	0.00
67 C	Ethyl Benzene	1.917	1.969	-2.7#	92	0.00
68 T	m/p-Xylenes	0.731	0.755	-3.3	92	0.00
69 T	o-Xylene	0.677	0.713	-5.3	92	0.00
70 T	Styrene	1.135	1.206	-6.3	93	0.00
71 P	Bromoform	0.254	0.247	2.8	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.695	3.828	-3.6	96	0.00
74 T	N-amyl acetate	1.215	1.033	15.0	71	0.00
75 P	1,1,2,2-Tetrachloroethane	0.831	0.749	9.9	81	0.00
76 T	1,2,3-Trichloropropane	0.623	0.597	4.2	86	0.00
77 T	Bromobenzene	0.885	0.888	-0.3	93	0.00
78 T	n-propylbenzene	4.622	4.695	-1.6	95	0.00
79 T	2-Chlorotoluene	2.568	2.559	0.4	94	0.00
80 T	1,3,5-Trimethylbenzene	3.086	3.148	-2.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.294	0.264	10.2	79	0.00
82 T	4-Chlorotoluene	2.663	2.649	0.5	94	0.00
83 T	tert-Butylbenzene	2.636	2.721	-3.2	95	0.00
84 T	1,2,4-Trimethylbenzene	3.017	3.075	-1.9	94	0.00
85 T	sec-Butylbenzene	4.010	4.156	-3.6	97	0.00
86 T	p-Isopropyltoluene	3.303	3.422	-3.6	97	0.00
87 T	1,3-Dichlorobenzene	1.734	1.752	-1.0	95	0.00
88 T	1,4-Dichlorobenzene	1.758	1.715	2.4	93	0.00
89 T	n-Butylbenzene	3.163	3.275	-3.5	97	0.00
90 T	Hexachloroethane	0.670	0.695	-3.7	101	0.00
91 T	1,2-Dichlorobenzene	1.549	1.541	0.5	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.137	0.111	19.0	72	0.00
93 T	1,2,4-Trichlorobenzene	0.979	0.959	2.0	92	0.00
94 T	Hexachlorobutadiene	0.591	0.582	1.5	97	0.00
95 T	Naphthalene	1.979	1.824	7.8	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
Data File : VY012750.D
Acq On : 01 Mar 2023 16:56
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 02 02:05:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.867	0.826	4.7	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6