

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012745.D
 Acq On : 01 Mar 2023 12:43
 Operator : KP/MD
 Sample : 01728-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-022723

Quant Time: Mar 02 02:04:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 22 01:21:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	137035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	225934	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	191905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	86040	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	94645	58.479	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.960%
35) Dibromofluoromethane	7.716	113	91345	59.858	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.720%
50) Toluene-d8	10.179	98	313010	54.749	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.500%
62) 4-Bromofluorobenzene	12.483	95	97965	53.546	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.100%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
Data File : VY012745.D
Acq On : 01 Mar 2023 12:43
Operator : KP/MD
Sample : 01728-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-022723

Quant Time: Mar 02 02:04:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
Quant Title : SW846 8260
QLast Update : Wed Feb 22 01:21:05 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43




Instrument :
MSVOA_Y
ClientSampleId :
OR-03-022723


Tgt Ion:168 Resp: 137035
Ion Ratio Lower Upper
168 100
99 52.3 45.4 68.2



#33
1,2-Dichloroethane-d4
Concen: 58.479 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. -0.001 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43

Tgt Ion: 65 Resp: 94645
Ion Ratio Lower Upper
65 100
67 52.3 0.0 104.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. -0.000 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43




Instrument :
MSVOA_Y
ClientSampleId :
OR-03-022723


Tgt Ion:114 Resp: 225934
Ion Ratio Lower Upper
114 100
63 19.9 0.0 42.8
88 14.4 0.0 29.6



#35
Dibromofluoromethane
Concen: 59.858 ug/l
RT: 7.716 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43

Tgt Ion:113 Resp: 91345
Ion Ratio Lower Upper
113 100
111 102.1 82.0 123.0
192 20.8 16.2 24.2









#50
Toluene-d8
Concen: 54.749 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43




Instrument :
MSVOA_Y
ClientSampleId :
OR-03-022723


Tgt Ion: 98 Resp: 313010
Ion Ratio Lower Upper
98 100
100 63.4 51.2 76.8



#62
4-Bromofluorobenzene
Concen: 53.546 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43

Tgt Ion: 95 Resp: 97965
Ion Ratio Lower Upper
95 100
174 88.8 0.0 172.8
176 85.4 0.0 169.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-022723

Tgt Ion:117 Resp: 191905
Ion Ratio Lower Upper
117 100
82 53.2 44.1 66.1
119 32.2 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY012745.D
Acq: 01 Mar 2023 12:43

Tgt Ion:152 Resp: 86040
Ion Ratio Lower Upper
152 100
115 56.3 28.7 86.3
150 155.0 0.0 347.6

