



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030123\
 Data File : VY012740.D
 Acq On : 01 Mar 2023 09:11
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022123S.M
 Title : SW846 8260
 Last Update : Wed Feb 22 01:21:05 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	11033	PASS
75	95	30	60	52.6	27683	PASS
95	95	100	100	100.0	52667	PASS
96	95	5	9	6.8	3597	PASS
173	174	0.00	2	1.7	762	PASS
174	95	50	100	86.7	45688	PASS
175	174	5	9	7.8	3568	PASS
176	174	95	101	95.0	43397	PASS
177	176	5	9	6.6	2865	PASS