

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
 Data File : VY012770.D
 Acq On : 02 Mar 2023 19:39
 Operator : KP/MD
 Sample : 01739-04
 Misc : 9.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-1-GRAB-7

Quant Time: Mar 03 01:59:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	129453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	216307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	169514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	64197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	76119	59.223	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.440%
35) Dibromofluoromethane	7.716	113	75258	57.651	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.300%
50) Toluene-d8	10.179	98	230916	48.387	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.483	95	69590	43.203	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
Data File : VY012770.D
Acq On : 02 Mar 2023 19:39
Operator : KP/MD
Sample : 01739-04
Misc : 9.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-1-GRAB-7

Quant Time: Mar 03 01:59:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

