

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
 Data File : VY012775.D
 Acq On : 02 Mar 2023 21:32
 Operator : KP/MD
 Sample : 01767-01
 Misc : 4.53g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CE-45-030123

Quant Time: Mar 03 02:01:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	119994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	200521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	159748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	63712	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	74716	62.714	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.420%
35) Dibromofluoromethane	7.716	113	70587	58.330	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.660%
50) Toluene-d8	10.179	98	217243	49.105	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.220%
62) 4-Bromofluorobenzene	12.483	95	65188	43.656	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
Data File : VY012775.D
Acq On : 02 Mar 2023 21:32
Operator : KP/MD
Sample : 01767-01
Misc : 4.53g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CE-45-030123

Quant Time: Mar 03 02:01:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

