

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
 Data File : VY012778.D
 Acq On : 02 Mar 2023 22:40
 Operator : KP/MD
 Sample : 01767-13
 Misc : 6.30g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CE-33-030123

Quant Time: Mar 03 02:02:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	118402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	195026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	154348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	56009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	70889	60.301	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.600%
35) Dibromofluoromethane	7.716	113	68224	57.965	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.940%
50) Toluene-d8	10.179	98	224110	52.085	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.160%
62) 4-Bromofluorobenzene	12.483	95	60716	41.807	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.620%
Target Compounds						
16) Acetone	3.948	43	9242	20.029	ug/l	96
95) Naphthalene	15.239	128	14880	7.173	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
Data File : VY012778.D
Acq On : 02 Mar 2023 22:40
Operator : KP/MD
Sample : 01767-13
Misc : 6.30g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CE-33-030123

Quant Time: Mar 03 02:02:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

