

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
 Data File : VY012779.D
 Acq On : 02 Mar 2023 23:03
 Operator : KP/MD
 Sample : 01760-03
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11998

Quant Time: Mar 03 02:02:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	1585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	1176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	170	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	737	83.591	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	167.180%#
35) Dibromofluoromethane	7.722	113	617	64.503	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.000%
50) Toluene-d8	10.179	98	1651	47.213	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.420%
62) 4-Bromofluorobenzene	12.495	95	366	31.009	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	=	62.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030223\
Data File : VY012779.D
Acq On : 02 Mar 2023 23:03
Operator : KP/MD
Sample : 01760-03
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11998

Quant Time: Mar 03 02:02:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

