


Dichlorodifluoromethane

Response Ratio



$$\text{Response} = 3.043\text{e-}001 * \text{Amt} + 2.909\text{e-}002$$


Coef of Det (r^2) = 0.990664 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y030321S.M

Calibration Table Last Updated: Wed Mar 03 13:49:02 2021

Chloromethane

Response Ratio



$$\text{Response} = 3.754\text{e-}001 * \text{Amt} + 2.251\text{e-}002$$


Coef of Det (r^2) = 0.993240 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y030321S.M

Calibration Table Last Updated: Wed Mar 03 13:49:02 2021

Vinyl Chloride


Response Ratio



Response = 4.410e-001 * Amt + 2.846e-002
 Coef of Det (r^2) = 0.995739 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y030321S.M
 Calibration Table Last Updated: Wed Mar 03 13:49:02 2021

Chloroethane

Response Ratio



$$\text{Response} = 2.808\text{e-}001 * \text{Amt} + 1.854\text{e-}002$$


Coef of Det (r^2) = 0.995120 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y030321S.M

Calibration Table Last Updated: Wed Mar 03 13:49:02 2021

Methylene Chloride

Response Ratio



$$\text{Response} = 4.558\text{e-}001 * \text{Amt} + 3.314\text{e-}002$$


Coef of Det (r^2) = 0.990520 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y030321S.M

Calibration Table Last Updated: Wed Mar 03 13:49:02 2021

2-Chloroethyl Vinyl ether

Response Ratio



$R = -2.700e-003 A^2 + 1.739e-001 A - 8.194e-003$
 Coef of Det (r^2) = 0.994349 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y030321S.M
 Calibration Table Last Updated: Wed Mar 03 13:49:02 2021