

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030323\
 Data File : VY012802.D
 Acq On : 03 Mar 2023 17:34
 Operator : KP/MD
 Sample : 01749-08
 Misc : 6.72g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CF-10

Quant Time: Mar 04 01:25:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	111653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	184300	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	157971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71540	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	68801	62.063	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.120%
35) Dibromofluoromethane	7.716	113	66390	59.690	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.380%
50) Toluene-d8	10.179	98	223722	55.020	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.040%
62) 4-Bromofluorobenzene	12.483	95	69142	50.379	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030323\
Data File : VY012802.D
Acq On : 03 Mar 2023 17:34
Operator : KP/MD
Sample : 01749-08
Misc : 6.72g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CF-10

Quant Time: Mar 04 01:25:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

