

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030323\
 Data File : VY012803.D
 Acq On : 03 Mar 2023 17:57
 Operator : KP/MD
 Sample : 01749-09
 Misc : 6.61g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CF-11

Quant Time: Mar 04 01:26:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	113976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	188654	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	166301	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77489	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	74839	66.134	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.260%
35) Dibromofluoromethane	7.716	113	70351	61.791	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	123.580%
50) Toluene-d8	10.179	98	238612	57.328	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.660%
62) 4-Bromofluorobenzene	12.483	95	73526	52.337	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.680%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030323\
Data File : VY012803.D
Acq On : 03 Mar 2023 17:57
Operator : KP/MD
Sample : 01749-09
Misc : 6.61g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CF-11

Quant Time: Mar 04 01:26:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY012803.D
Acq: 03 Mar 2023 17:57




Instrument :
MSVOA_Y
ClientSampleId :
CF-11


Tgt Ion:168 Resp: 113976
Ion Ratio Lower Upper
168 100
99 54.1 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 66.134 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. -0.001 min
Lab File: VY012803.D
Acq: 03 Mar 2023 17:57

Tgt Ion: 65 Resp: 74839
Ion Ratio Lower Upper
65 100
67 51.6 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY012803.D

Acq: 03 Mar 2023 17:57

Instrument :

MSVOA_Y

ClientSampleId :

CF-11





Tgt Ion:114 Resp: 188654

Ion Ratio Lower Upper

114 100

63 19.6 0.0 37.4

88 15.6 0.0 28.6



#35

Dibromofluoromethane

Concen: 61.791 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY012803.D

Acq: 03 Mar 2023 17:57




Tgt Ion:113 Resp: 70351


Ion Ratio Lower Upper

113 100

111 102.0 82.2 123.2

192 20.9 16.8 25.2









#50
Toluene-d8
Concen: 57.328 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY012803.D
Acq: 03 Mar 2023 17:57




Instrument :
MSVOA_Y
ClientSampleId :
CF-11


Tgt Ion: 98 Resp: 238612
Ion Ratio Lower Upper
98 100
100 63.1 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 52.337 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY012803.D
Acq: 03 Mar 2023 17:57

Tgt Ion: 95 Resp: 73526
Ion Ratio Lower Upper
95 100
174 91.5 0.0 176.6
176 87.1 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY012803.D
Acq: 03 Mar 2023 17:57

Instrument :
MSVOA_Y
ClientSampleId :
CF-11

Tgt Ion:117 Resp: 166301
Ion Ratio Lower Upper
117 100
82 55.9 45.4 68.0
119 31.6 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY012803.D
Acq: 03 Mar 2023 17:57

Tgt Ion:152 Resp: 77489
Ion Ratio Lower Upper
152 100
115 55.3 27.1 81.3
150 158.1 0.0 347.8

