



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030323\
 Data File : VY012784.D
 Acq On : 03 Mar 2023 09:30
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Title : SW846 8260
 Last Update : Fri Mar 03 01:13:13 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	10730	PASS
75	95	30	60	53.3	27667	PASS
95	95	100	100	100.0	51939	PASS
96	95	5	9	6.3	3274	PASS
173	174	0.00	2	1.7	732	PASS
174	95	50	100	84.8	44035	PASS
175	174	5	9	7.9	3474	PASS
176	174	95	101	95.5	42037	PASS
177	176	5	9	6.9	2904	PASS