

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030524\
 Data File : VY017567.D
 Acq On : 05 Mar 2024 13:31
 Operator : SY/MD
 Sample : P1674-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 1A

Quant Time: Mar 05 22:03:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	97442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	141256	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	101540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	32206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	30389	32.907	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	65.820%
35) Dibromofluoromethane	7.716	113	40050	40.688	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.380%
50) Toluene-d8	10.185	98	153421	44.579	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.160%
62) 4-Bromofluorobenzene	12.489	95	30650	28.338	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	56.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030524\
Data File : VY017567.D
Acq On : 05 Mar 2024 13:31
Operator : SY/MD
Sample : P1674-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
1A

Quant Time: Mar 05 22:03:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

