

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030524\
 Data File : VY017569.D
 Acq On : 05 Mar 2024 14:19
 Operator : SY/MD
 Sample : P1669-03
 Misc : 3.09g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-14-VOC

Quant Time: Mar 05 22:04:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	111547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	201753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	182291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	70401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	63594	60.155	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.320%
35) Dibromofluoromethane	7.722	113	67330	47.892	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.780%
50) Toluene-d8	10.185	98	228562	46.498	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.000%
62) 4-Bromofluorobenzene	12.489	95	63652	41.204	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.400%
Target Compounds						
20) Methylene Chloride	4.723	84	20518	12.975	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030524\
Data File : VY017569.D
Acq On : 05 Mar 2024 14:19
Operator : SY/MD
Sample : P1669-03
Misc : 3.09g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-14-VOC

Quant Time: Mar 05 22:04:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

