

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030524\
 Data File : VY017571.D
 Acq On : 05 Mar 2024 15:07
 Operator : SY/MD
 Sample : P1672-11
 Misc : 3.97g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-5-VOC

Quant Time: Mar 05 22:05:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	111969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	199493	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	178294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	66397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	55593	52.389	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.780%
35) Dibromofluoromethane	7.728	113	64621	46.486	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.980%
50) Toluene-d8	10.191	98	225887	46.475	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.940%
62) 4-Bromofluorobenzene	12.489	95	61532	40.283	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.560%
Target Compounds						
20) Methylene Chloride	4.722	84	14647	7.451	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030524\
Data File : VY017571.D
Acq On : 05 Mar 2024 15:07
Operator : SY/MD
Sample : P1672-11
Misc : 3.97g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-5-VOC

Quant Time: Mar 05 22:05:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

