



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030624\
Data File : VY017579.D
Acq On : 06 Mar 2024 08:26
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Title : SW846 8260
Last Update : Sat Feb 24 04:41:20 2024



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.8	7377	PASS
75	95	30	60	47.8	20960	PASS
95	95	100	100	100.0	43880	PASS
96	95	5	9	6.6	2898	PASS
173	174	0.00	2	1.2	508	PASS
174	95	50	100	99.8	43792	PASS
175	174	5	9	7.7	3386	PASS
176	174	95	101	96.2	42117	PASS
177	176	5	9	6.7	2808	PASS