

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
 Data File : VY021440.D
 Acq On : 06 Mar 2025 17:22
 Operator : SY/MD
 Sample : Q1488-11
 Misc : 6.49g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ENV-104-SB02

Quant Time: Mar 07 00:41:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.707	168	948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	243	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	521	51.997	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.000%	
35) Dibromofluoromethane	7.646	113	539	83.845	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 167.680%#	
50) Toluene-d8	10.109	98	746	30.649	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 61.300%	
62) 4-Bromofluorobenzene	12.426	95	63	7.609	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	= 15.220%#	
Target Compounds						Qvalue
3) Chloromethane	2.196	50	46	3.751	ug/l #	42
4) Vinyl Chloride	2.013	62	33	2.462	ug/l #	42
6) Chloroethane	2.903	64	34	3.839	ug/l #	41
10) Methyl Iodide	3.994	142	41	3.776	ug/l #	37
11) Tert butyl alcohol	4.763	59	22	12.349	ug/l #	72
12) 1,1-Dichloroethene	3.885	96	51	5.237	ug/l #	1
13) Acrolein	3.805	56	21	41.205	ug/l #	10
14) Allyl chloride	4.385	41	36	2.288	ug/l #	21
16) Acetone	3.854	43	178	81.449	ug/l #	47
18) Methyl Acetate	4.311	43	26	5.362	ug/l #	51
19) Methyl tert-butyl Ether	4.884	73	29	1.175	ug/l #	53
21) trans-1,2-Dichloroethene	5.195	96	41	3.759	ug/l #	1
23) Vinyl Acetate	5.835	43	22	1.137	ug/l #	72
25) 2-Butanone	7.061	43	39	12.901	ug/l #	48
27) cis-1,2-Dichloroethene	6.732	96	29	2.341	ug/l #	72
28) Bromochloromethane	7.165	49	26	3.098	ug/l #	14
29) Tetrahydrofuran	7.238	42	23	12.324	ug/l #	33
31) Cyclohexane	7.927	56	31	1.652	ug/l #	15
36) 1,1-Dichloropropene	7.707	75	100	9.938	ug/l #	41
37) Ethyl Acetate	7.061	43	39	8.769	ug/l #	68
41) Methacrylonitrile	7.232	41	46	18.995	ug/l #	24
42) 1,2-Dichloroethane	8.335	62	19	2.291	ug/l #	75
43) Isopropyl Acetate	8.231	43	25	2.896	ug/l #	52
47) Bromodichloromethane	9.609	83	21	2.014	ug/l #	25
48) Methyl methacrylate	9.219	41	47	11.855	ug/l #	1
49) 1,4-Dioxane	9.128	88	45	1167.942	ug/l #	22
51) 4-Methyl-2-Pentanone	10.152	43	54	12.122	ug/l #	36
55) 1,1,2-Trichloroethane	10.774	97	20	3.895	ug/l #	14
56) Ethyl methacrylate	10.518	69	20	3.054	ug/l #	24
59) 2-Hexanone	10.871	43	29	9.845	ug/l #	24


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
Data File : VY021440.D
Acq On : 06 Mar 2025 17:22
Operator : SY/MD
Sample : Q1488-11
Misc : 6.49g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02

Quant Time: Mar 07 00:41:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22




Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02


Tgt Ion:168 Resp: 948
Ion Ratio Lower Upper
168 100
99 53.2 46.0 69.0



#3
Chloromethane
Concen: 3.751 ug/l
RT: 2.196 min Scan# 78
Delta R.T. 0.122 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion: 50 Resp: 46
Ion Ratio Lower Upper
50 100
52 0.0 26.2 39.2#









#4
 Vinyl Chloride
 Concen: 2.462 ug/l
 RT: 2.013 min Scan# 41
 Delta R.T. -0.195 min
 Lab File: VY021440.D
 Acq: 06 Mar 2025 17:22




Instrument :
 MSVOA_Y
 ClientSampleId :
 ENV-104-SB02


Tgt Ion: 62 Resp: 33
 Ion Ratio Lower Upper
 62 100
 64 0.0 26.2 39.2#



#6
 Chloroethane
 Concen: 3.839 ug/l
 RT: 2.903 min Scan# 194
 Delta R.T. 0.165 min
 Lab File: VY021440.D
 Acq: 06 Mar 2025 17:22


Tgt Ion: 64 Resp: 34
 Ion Ratio Lower Upper
 64 100
 66 0.0 26.7 40.1#








#10
Methyl Iodide
Concen: 3.776 ug/l
RT: 3.994 min Scan# 31
Delta R.T. -0.018 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02







Tgt Ion:142 Resp: 41
Ion Ratio Lower Upper
142 100
127 0.0 39.1 58.7#
141 0.0 11.8 17.6#



#11
Tert butyl alcohol
Concen: 12.349 ug/l
RT: 4.763 min Scan# 499
Delta R.T. -0.109 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion: 59 Resp: 22
Ion Ratio Lower Upper
59 100
57 0.0 8.2 12.4#





#12

1,1-Dichloroethene

Concen: 5.237 ug/l

RT: 3.885 min Scan# 31

Delta R.T. 0.086 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Instrument :

MSVOA_Y

ClientSampleId :

ENV-104-SB02





Tgt Ion: 96 Resp: 51

Ion Ratio Lower Upper

96 100

61 0.0 135.2 202.8#

98 0.0 51.4 77.2#



#13

Acrolein

Concen: 41.205 ug/l

RT: 3.805 min Scan# 342

Delta R.T. 0.140 min

Lab File: VY021440.D




Acq: 06 Mar 2025 17:22


Tgt Ion: 56 Resp: 21

Ion Ratio Lower Upper

56 100

55 0.0 63.4 95.2#





#14

Allyl chloride

Concen: 2.288 ug/l

RT: 4.385 min Scan# 41

Delta R.T. -0.012 min

Lab File: VY021440.D


Acq: 06 Mar 2025 17:22

Instrument :




MSVOA_Y

ClientSampleId :

ENV-104-SB02



Tgt Ion:	41	Resp:	36
Ion	Ratio	Lower	Upper
41	100		
39	0.0	59.3	88.9#
76	0.0	29.8	44.8#



#16

Acetone

Concen: 81.449 ug/l




RT: 3.854 min Scan# 350


Delta R.T. -0.018 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Tgt Ion:	43	Resp:	178
Ion	Ratio	Lower	Upper
43	100		
58	0.0	22.1	33.1#





#18

Methyl Acetate

Concen: 5.362 ug/l

RT: 4.311 min Scan# 41

Delta R.T. -0.079 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Instrument :

MSVOA_Y

ClientSampleId :





ENV-104-SB02

Tgt Ion: 43 Resp: 26

Ion Ratio Lower Upper

43 100

74 0.0 19.4 29.0#



#19

Methyl tert-butyl Ether

Concen: 1.175 ug/l

RT: 4.884 min Scan# 519

Delta R.T. -0.238 min

Lab File: VY021440.D




Acq: 06 Mar 2025 17:22


Tgt Ion: 73 Resp: 29

Ion Ratio Lower Upper

73 100

57 0.0 17.9 26.9#





#21

trans-1,2-Dichloroethene

Concen: 3.759 ug/l

RT: 5.195 min Scan# 51

Delta R.T. 0.067 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Instrument :

MSVOA_Y

ClientSampleId :

ENV-104-SB02





Tgt Ion: 96 Resp: 41

Ion Ratio Lower Upper

96 100

61 0.0 111.5 167.3#

98 0.0 53.8 80.8#



#23

Vinyl Acetate

Concen: 1.137 ug/l

RT: 5.835 min Scan# 675

Delta R.T. -0.134 min

Lab File: VY021440.D




Acq: 06 Mar 2025 17:22


Tgt Ion: 43 Resp: 22

Ion Ratio Lower Upper

43 100

86 0.0 8.5 12.7#





#25

2-Butanone

Concen: 12.901 ug/l

RT: 7.061 min Scan# 81

Delta R.T. 0.165 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Instrument :

MSVOA_Y

ClientSampleId :





ENV-104-SB02

Tgt Ion: 43 Resp: 39

Ion Ratio Lower Upper

43 100

72 0.0 21.4 32.0#



#27

cis-1,2-Dichloroethene

Concen: 2.341 ug/l

RT: 6.732 min Scan# 822

Delta R.T. -0.164 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22




Tgt Ion: 96 Resp: 29


Ion Ratio Lower Upper

96 100

61 144.8 0.0 300.8

98 0.0 0.0 130.0









#28
Bromochloromethane
Concen: 3.098 ug/l
RT: 7.165 min Scan# 893
Delta R.T. -0.085 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22




Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02


Tgt Ion: 49 Resp: 26
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.2
130 0.0 59.6 89.4#



#29
Tetrahydrofuran
Concen: 12.324 ug/l
RT: 7.238 min Scan# 905
Delta R.T. -0.030 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion: 42 Resp: 23
Ion Ratio Lower Upper
42 100
72 0.0 34.8 52.2#
71 0.0 32.6 48.8#









#31
Cyclohexane
Concen: 1.652 ug/l
RT: 7.927 min Scan# 1018
Delta R.T. 0.220 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22




Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02


Tgt Ion: 56 Resp: 31
Ion Ratio Lower Upper
56 100
69 0.0 27.0 40.4#
84 0.0 71.1 106.7#



#33
1,2-Dichloroethane-d4
Concen: 51.997 ug/l
RT: 8.055 min Scan# 1039
Delta R.T. -0.012 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion: 65 Resp: 521
Ion Ratio Lower Upper
65 100
67 33.6 0.0 102.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.622 min Scan# 1132

Delta R.T. 0.000 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22





Instrument :

MSVOA_Y

ClientSampleId :

ENV-104-SB02

Tgt Ion:114	Resp:	978
Ion	Ratio	Lower Upper
114	100	
63	21.9	0.0 40.8
88	13.0	0.0 30.8



#35

Dibromofluoromethane

Concen: 83.845 ug/l




RT: 7.646 min Scan# 972


Delta R.T. 0.006 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Tgt Ion:113	Resp:	539
Ion	Ratio	Lower Upper
113	100	
111	96.1	82.0 123.0
192	11.3	15.9 23.9









#36
 1,1-Dichloropropene
 Concen: 9.938 ug/l
 RT: 7.707 min Scan# 91
 Delta R.T. -0.134 min
 Lab File: VY021440.D
 Acq: 06 Mar 2025 17:22




Instrument :
 MSVOA_Y
 ClientSampleId :
 ENV-104-SB02


Tgt Ion: 75 Resp: 100
 Ion Ratio Lower Upper
 75 100
 110 0.0 17.3 51.7#
 77 0.0 25.4 38.2#



#37
 Ethyl Acetate
 Concen: 8.769 ug/l
 RT: 7.061 min Scan# 876
 Delta R.T. 0.073 min
 Lab File: VY021440.D
 Acq: 06 Mar 2025 17:22


Tgt Ion: 43 Resp: 39
 Ion Ratio Lower Upper
 43 100
 61 0.0 11.4 17.0#
 70 0.0 8.2 12.2#








#41
Methacrylonitrile
Concen: 18.995 ug/l
RT: 7.232 min Scan# 904
Delta R.T. 0.012 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02







Tgt Ion	Ratio	Lower	Upper
41	100		
39	0.0	35.0	52.6#
67	0.0	61.2	91.8#
52	0.0	26.8	40.2#



#42
1,2-Dichloroethane
Concen: 2.291 ug/l
RT: 8.335 min Scan# 1085
Delta R.T. 0.171 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	18.2









#43
Isopropyl Acetate
Concen: 2.896 ug/l
RT: 8.231 min Scan# 1063
Delta R.T. 0.031 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22




Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02


Tgt Ion: 43 Resp: 25
Ion Ratio Lower Upper
43 100
61 0.0 23.4 35.0#
87 0.0 10.5 15.7#



#47
Bromodichloromethane
Concen: 2.014 ug/l
RT: 9.609 min Scan# 1294
Delta R.T. 0.183 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion: 83 Resp: 21
Ion Ratio Lower Upper
83 100
85 0.0 51.1 76.7#
127 0.0 6.7 10.1#





#48

Methyl methacrylate

Concen: 11.855 ug/l

RT: 9.219 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Instrument :

MSVOA_Y

ClientSampleId :

ENV-104-SB02





Tgt Ion: 41 Resp: 47

Ion Ratio Lower Upper

41 100

69 0.0 67.8 101.8#

39 180.9 46.4 69.6#



#49

1,4-Dioxane

Concen: 1167.942 ug/l

RT: 9.128 min Scan# 1215

Delta R.T. -0.103 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22




Tgt Ion: 88 Resp: 45


Ion Ratio Lower Upper

88 100

43 0.0 25.0 37.4#

58 0.0 58.4 87.6#





#50

Toluene-d8

Concen: 30.649 ug/l

RT: 10.109 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Instrument :

MSVOA_Y

ClientSampleId :





ENV-104-SB02

Tgt Ion: 98 Resp: 746

Ion Ratio Lower Upper

98 100

100 61.8 52.1 78.1



#51

4-Methyl-2-Pentanone

Concen: 12.122 ug/l

RT: 10.152 min Scan# 1383

Delta R.T. 0.153 min

Lab File: VY021440.D




Acq: 06 Mar 2025 17:22


Tgt Ion: 43 Resp: 54

Ion Ratio Lower Upper

43 100

58 0.0 31.4 47.2#





#55

1,1,2-Trichloroethane

Concen: 3.895 ug/l

RT: 10.774 min Scan# 1445

Delta R.T. 0.201 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22

Instrument :

MSVOA_Y

ClientSampleId :

ENV-104-SB02

Tgt Ion: 97 Resp: 20





Ion Ratio Lower Upper

97 100

83 0.0 70.0 105.0#

85 0.0 44.3 66.5#

99 0.0 51.4 77.0#



#56

Ethyl methacrylate

Concen: 3.054 ug/l

RT: 10.518 min Scan# 1443

Delta R.T. 0.073 min

Lab File: VY021440.D

Acq: 06 Mar 2025 17:22




Tgt Ion: 69 Resp: 20


Ion Ratio Lower Upper

69 100

41 0.0 53.3 79.9#

39 0.0 31.0 46.6#









#59
2-Hexanone
Concen: 9.845 ug/l
RT: 10.871 min Scan# 11
Delta R.T. 0.110 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22




Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02


Tgt Ion: 43 Resp: 29
Ion Ratio Lower Upper
43 100
58 0.0 27.3 81.8#



#62
4-Bromofluorobenzene
Concen: 7.609 ug/l
RT: 12.426 min Scan# 1756
Delta R.T. 0.018 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion: 95 Resp: 63
Ion Ratio Lower Upper
95 100
174 0.0 0.0 165.0
176 0.0 0.0 160.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Instrument :
MSVOA_Y
ClientSampleId :
ENV-104-SB02

Tgt Ion:117 Resp: 243
Ion Ratio Lower Upper
117 100
82 66.4 44.6 67.0
119 0.0 25.4 38.0#



#72
1,4-Dichlorobenzene-d4
Concen: 0.000 ug/l
Expected RT: 13.35 min

Lab File: VY021440.D
Acq: 06 Mar 2025 17:22

Tgt Ion: 152
Sig Exp Ratio
152 100
115 58.0
150 173.6

