

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
 Data File : VY021437.D
 Acq On : 06 Mar 2025 16:12
 Operator : SY/MD
 Sample : Q1488-05
 Misc : 7.37g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ENV-102-SB01

Quant Time: Mar 07 00:41:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	245112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	450789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	408673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	161699	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	165897	64.036	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.080%
35) Dibromofluoromethane	7.634	113	156243	52.730	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.460%
50) Toluene-d8	10.103	98	552261	49.224	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.440%
62) 4-Bromofluorobenzene	12.408	95	176528	46.256	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
Data File : VY021437.D
Acq On : 06 Mar 2025 16:12
Operator : SY/MD
Sample : Q1488-05
Misc : 7.37g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ENV-102-SB01

Quant Time: Mar 07 00:41:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

