

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
 Data File : VY021447.D
 Acq On : 06 Mar 2025 20:06
 Operator : SY/MD
 Sample : Q1463-09
 Misc : 6.48g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T9

Quant Time: Mar 07 00:43:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	211507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	387107	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	337914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	129134	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	136932	61.253	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.500%
35) Dibromofluoromethane	7.634	113	130951	51.464	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.920%
50) Toluene-d8	10.103	98	457447	47.481	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.960%
62) 4-Bromofluorobenzene	12.408	95	137652	42.003	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
Data File : VY021447.D
Acq On : 06 Mar 2025 20:06
Operator : SY/MD
Sample : Q1463-09
Misc : 6.48g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T9

Quant Time: Mar 07 00:43:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

