

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
 Data File : VY021441.D
 Acq On : 06 Mar 2025 17:46
 Operator : SY/MD
 Sample : Q1463-01
 Misc : 6.09g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T1

Quant Time: Mar 07 00:41:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	222165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	418468	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	361430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	139247	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	148589	63.279	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.560%
35) Dibromofluoromethane	7.634	113	144278	52.452	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.900%
50) Toluene-d8	10.109	98	419298	40.260	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	80.520%
62) 4-Bromofluorobenzene	12.408	95	149255	42.130	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.260%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030625\
Data File : VY021441.D
Acq On : 06 Mar 2025 17:46
Operator : SY/MD
Sample : Q1463-01
Misc : 6.09g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T1

Quant Time: Mar 07 00:41:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 99
Delta R.T. -0.006 min
Lab File: VY021441.D
Acq: 06 Mar 2025 17:46




Instrument :
MSVOA_Y
ClientSampleId :
T1


Tgt Ion:168 Resp: 222165
Ion Ratio Lower Upper
168 100
99 56.6 46.0 69.0



#33
1,2-Dichloroethane-d4
Concen: 63.279 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VY021441.D
Acq: 06 Mar 2025 17:46

Tgt Ion: 65 Resp: 148589
Ion Ratio Lower Upper
65 100
67 53.1 0.0 102.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1132

Delta R.T. -0.006 min

Lab File: VY021441.D

Acq: 06 Mar 2025 17:46

Instrument :

MSVOA_Y

ClientSampleId :

T1





Tgt Ion:114 Resp: 418468

Ion Ratio Lower Upper

114 100

63 21.3 0.0 40.8

88 15.2 0.0 30.8



#35

Dibromofluoromethane

Concen: 52.452 ug/l

RT: 7.634 min Scan# 970

Delta R.T. -0.006 min

Lab File: VY021441.D

Acq: 06 Mar 2025 17:46




Tgt Ion:113 Resp: 144278


Ion Ratio Lower Upper

113 100

111 102.5 82.0 123.0

192 19.0 15.9 23.9









#50
Toluene-d8
Concen: 40.260 ug/l
RT: 10.109 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY021441.D
Acq: 06 Mar 2025 17:46




Instrument :
MSVOA_Y
ClientSampleId :
T1


Tgt Ion: 98 Resp: 419298
Ion Ratio Lower Upper
98 100
100 64.4 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 42.130 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY021441.D
Acq: 06 Mar 2025 17:46

Tgt Ion: 95 Resp: 149255
Ion Ratio Lower Upper
95 100
174 81.2 0.0 165.0
176 78.1 0.0 160.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY021441.D
Acq: 06 Mar 2025 17:46

Instrument :
MSVOA_Y
ClientSampleId :
T1

Tgt Ion:117 Resp: 361430
Ion Ratio Lower Upper
117 100
82 57.0 44.6 67.0
119 32.3 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.347 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VY021441.D
Acq: 06 Mar 2025 17:46

Tgt Ion:152 Resp: 139247
Ion Ratio Lower Upper
152 100
115 59.4 29.0 87.0
150 156.6 0.0 347.2

