

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030723\
 Data File : VY012851.D
 Acq On : 07 Mar 2023 17:42
 Operator : KP/MD
 Sample : 01820-01RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-3-03032023RE

Quant Time: Mar 07 23:44:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	299	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	761	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	639	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	496	167.077	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	334.160%#
35) Dibromofluoromethane	7.716	113	224	48.774	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.540%
50) Toluene-d8	10.179	98	866	51.579	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.160%
62) 4-Bromofluorobenzene	12.489	95	114	20.117	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	40.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030723\
Data File : VY012851.D
Acq On : 07 Mar 2023 17:42
Operator : KP/MD
Sample : 01820-01RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-3-03032023RE

Quant Time: Mar 07 23:44:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

