

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030723\
 Data File : VY012856.D
 Acq On : 07 Mar 2023 19:39
 Operator : KP/MD
 Sample : 01817-01RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PL-04-03032023RE

Quant Time: Mar 07 23:45:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	2705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	4014	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	3687	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1181	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	2238	83.330	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 166.660%#	
35) Dibromofluoromethane	7.722	113	1697	70.053	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 140.100%	
50) Toluene-d8	10.179	98	4839	54.641	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.280%	
62) 4-Bromofluorobenzene	12.489	95	1239	41.450	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 82.900%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030723\
Data File : VY012856.D
Acq On : 07 Mar 2023 19:39
Operator : KP/MD
Sample : 01817-01RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PL-04-03032023RE

Quant Time: Mar 07 23:45:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

