

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030724\
 Data File : VY017610.D
 Acq On : 07 Mar 2024 18:20
 Operator : SY/MD
 Sample : P1705-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Mar 08 03:28:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	47178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	74352	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	59612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	20227	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	20893	46.728	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.460%
35) Dibromofluoromethane	7.716	113	24671	47.617	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.240%
50) Toluene-d8	10.185	98	81353	44.909	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.820%
62) 4-Bromofluorobenzene	12.489	95	17934	31.502	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	63.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030724\
Data File : VY017610.D
Acq On : 07 Mar 2024 18:20
Operator : SY/MD
Sample : P1705-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Mar 08 03:28:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

