

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030724\
 Data File : VY017599.D
 Acq On : 07 Mar 2024 11:10
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 08 03:23:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.369	0.361	2.2	87	0.00
3 P	Chloromethane	0.674	0.667	1.0	88	0.00
4 C	Vinyl Chloride	0.865	0.914	-5.7#	92	0.00
5 T	Bromomethane	0.705	0.739	-4.8	94	0.00
6 T	Chloroethane	0.545	0.592	-8.6	92	0.00
7 T	Trichlorofluoromethane	0.881	0.925	-5.0	91	0.00
8 T	Diethyl Ether	0.258	0.272	-5.4	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.507	0.544	-7.3	92	0.00
10 T	Methyl Iodide	0.569	0.656	-15.3	92	0.00
11 T	Tert butyl alcohol	0.040	0.033	17.5	77	0.00
12 CM	1,1-Dichloroethene	0.472	0.513	-8.7#	93	0.00
13 T	Acrolein	0.053	0.051	3.8	83	0.00
14 T	Allyl chloride	0.571	0.607	-6.3	88	0.00
15 T	Acrylonitrile	0.106	0.109	-2.8	83	0.00
16 T	Acetone	0.092	0.102	-10.9	92	0.00
17 T	Carbon Disulfide	1.418	1.562	-10.2	93	0.00
18 T	Methyl Acetate	0.223	0.244	-9.4	86	0.00
19 T	Methyl tert-butyl Ether	1.101	1.166	-5.9	84	0.00
20 T	Methylene Chloride	0.658	0.630	4.3	95	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.609	-10.1	92	0.00
22 T	Diisopropyl ether	1.317	1.436	-9.0	87	0.00
23 T	Vinyl Acetate	0.785	0.846	-7.8	85	0.00
24 P	1,1-Dichloroethane	0.900	0.966	-7.3	90	0.00
25 T	2-Butanone	0.131	0.139	-6.1	85	0.00
26 T	2,2-Dichloropropane	0.756	0.843	-11.5	95	0.00
27 T	cis-1,2-Dichloroethene	0.621	0.675	-8.7	89	0.00
28 T	Bromochloromethane	0.354	0.364	-2.8	103	0.00
29 T	Tetrahydrofuran	0.079	0.082	-3.8	81	0.00
30 C	Chloroform	0.958	1.028	-7.3#	90	0.00
31 T	Cyclohexane	0.771	0.820	-6.4	92	0.00
32 T	1,1,1-Trichloroethane	0.832	0.908	-9.1	92	0.00
33 S	1,2-Dichloroethane-d4	0.474	0.456	3.8	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.348	0.353	-1.4	87	0.00
36 T	1,1-Dichloropropene	0.470	0.538	-14.5	92	0.00
37 T	Ethyl Acetate	0.186	0.194	-4.3	83	0.00
38 T	Carbon Tetrachloride	0.471	0.544	-15.5	92	0.00
39 T	Methylcyclohexane	0.565	0.671	-18.8	93	0.00
40 TM	Benzene	1.396	1.577	-13.0	91	0.00
41 T	Methacrylonitrile	0.093	0.107	-15.1	86	-0.02
42 TM	1,2-Dichloroethane	0.341	0.373	-9.4	87	0.00
43 T	Isopropyl Acetate	0.326	0.346	-6.1	81	0.00
44 TM	Trichloroethene	0.402	0.449	-11.7	91	0.00
45 C	1,2-Dichloropropane	0.332	0.370	-11.4#	90	0.00
46 T	Dibromomethane	0.190	0.210	-10.5	88	0.00
47 T	Bromodichloromethane	0.462	0.514	-11.3	88	0.00
48 T	Methyl methacrylate	0.151	0.167	-10.6	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030724\
 Data File : VY017599.D
 Acq On : 07 Mar 2024 11:10
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 08 03:23:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	81	0.00
50 S	Toluene-d8	1.218	1.332	-9.4	89	0.00
51 T	4-Methyl-2-Pentanone	0.182	0.196	-7.7	82	0.00
52 CM	Toluene	0.880	1.014	-15.2#	92	0.00
53 T	t-1,3-Dichloropropene	0.428	0.487	-13.8	88	0.00
54 T	cis-1,3-Dichloropropene	0.521	0.589	-13.1	89	0.00
55 T	1,1,2-Trichloroethane	0.270	0.293	-8.5	87	0.00
56 T	Ethyl methacrylate	0.303	0.342	-12.9	82	0.00
57 T	1,3-Dichloropropane	0.439	0.483	-10.0	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.156	0.171	-9.6	99	0.00
59 T	2-Hexanone	0.125	0.142	-13.6	84	0.00
60 T	Dibromochloromethane	0.319	0.364	-14.1	89	0.00
61 T	1,2-Dibromoethane	0.249	0.280	-12.4	88	0.00
62 S	4-Bromofluorobenzene	0.383	0.426	-11.2	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.480	0.500	-4.2	86	0.00
65 PM	Chlorobenzene	1.061	1.204	-13.5	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.452	-13.6	92	0.00
67 C	Ethyl Benzene	1.861	2.148	-15.4#	91	0.00
68 T	m/p-Xylenes	0.731	0.861	-17.8	93	0.00
69 T	o-Xylene	0.679	0.794	-16.9	90	0.00
70 T	Styrene	1.126	1.316	-16.9	89	0.00
71 P	Bromoform	0.220	0.249	-13.2	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.715	4.050	-9.0	91	0.00
74 T	N-amyl acetate	0.655	0.667	-1.8	81	0.00
75 P	1,1,2,2-Tetrachloroethane	0.638	0.695	-8.9	91	0.00
76 T	1,2,3-Trichloropropane	0.475	0.488	-2.7	90	0.00
77 T	Bromobenzene	0.900	0.957	-6.3	91	0.00
78 T	n-propylbenzene	4.457	4.930	-10.6	92	0.00
79 T	2-Chlorotoluene	2.495	2.631	-5.5	90	0.00
80 T	1,3,5-Trimethylbenzene	2.976	3.313	-11.3	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.207	0.217	-4.8	87	0.00
82 T	4-Chlorotoluene	2.552	2.732	-7.1	91	0.00
83 T	tert-Butylbenzene	2.613	2.909	-11.3	92	0.00
84 T	1,2,4-Trimethylbenzene	2.938	3.267	-11.2	92	0.00
85 T	sec-Butylbenzene	3.996	4.473	-11.9	92	0.00
86 T	p-Isopropyltoluene	3.271	3.738	-14.3	93	0.00
87 T	1,3-Dichlorobenzene	1.806	1.965	-8.8	93	0.00
88 T	1,4-Dichlorobenzene	1.755	1.882	-7.2	91	0.00
89 T	n-Butylbenzene	2.984	3.381	-13.3	92	0.00
90 T	Hexachloroethane	0.656	0.716	-9.1	94	0.00
91 T	1,2-Dichlorobenzene	1.527	1.681	-10.1	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.090	0.088	2.2	84	0.00
93 T	1,2,4-Trichlorobenzene	0.807	0.929	-15.1	93	0.00
94 T	Hexachlorobutadiene	0.498	0.578	-16.1	100	0.00
95 T	Naphthalene	1.379	1.534	-11.2	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030724\
Data File : VY017599.D
Acq On : 07 Mar 2024 11:10
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 08 03:23:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.692	0.777	-12.3	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6