

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030824\
 Data File : VY017614.D
 Acq On : 08 Mar 2024 09:51
 Operator : SY/MD
 Sample : VY0308SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0308SBL01

Quant Time: Mar 08 22:20:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	130340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	234810	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	211243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	80020	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	62773	50.817	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.640%
35) Dibromofluoromethane	7.728	113	75852	46.358	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.720%
50) Toluene-d8	10.185	98	269143	47.046	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.100%
62) 4-Bromofluorobenzene	12.483	95	72553	40.354	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030824\
Data File : VY017614.D
Acq On : 08 Mar 2024 09:51
Operator : SY/MD
Sample : VY0308SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0308SBL01

Quant Time: Mar 08 22:20:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

