

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030824\
 Data File : VY017621.D
 Acq On : 08 Mar 2024 13:22
 Operator : SY/MD
 Sample : P1710-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-030724

Quant Time: Mar 08 22:24:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	112395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	207458	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	184609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71841	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	56163	52.725	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.460%
35) Dibromofluoromethane	7.722	113	67343	46.584	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.160%
50) Toluene-d8	10.179	98	238753	47.236	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.480%
62) 4-Bromofluorobenzene	12.483	95	64545	40.633	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030824\
Data File : VY017621.D
Acq On : 08 Mar 2024 13:22
Operator : SY/MD
Sample : P1710-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-030724

Quant Time: Mar 08 22:24:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

