

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030824\
 Data File : VY017625.D
 Acq On : 08 Mar 2024 17:11
 Operator : SY/MD
 Sample : P1720-03RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT3471RE

Quant Time: Mar 08 23:39:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	69699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	123001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	114109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	46157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	35320	53.470	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.940%	
35) Dibromofluoromethane	7.716	113	39621	46.226	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 92.460%	
50) Toluene-d8	10.179	98	154271	51.479	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.960%	
62) 4-Bromofluorobenzene	12.483	95	44607	47.363	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.720%	
Target Compounds						
					Qvalue	
52) Toluene	10.240	92	3391	1.566	ug/l	93
73) Isopropylbenzene	12.337	105	45980	13.409	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030824\
Data File : VY017625.D
Acq On : 08 Mar 2024 17:11
Operator : SY/MD
Sample : P1720-03RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT3471RE

Quant Time: Mar 08 23:39:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

