



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY030824\
Data File : VY017612.D
Acq On : 08 Mar 2024 08:19
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Title : SW846 8260
Last Update : Sat Feb 24 04:41:20 2024



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	8093	PASS
75	95	30	60	50.4	23485	PASS
95	95	100	100	100.0	46608	PASS
96	95	5	9	6.7	3117	PASS
173	174	0.00	2	1.1	517	PASS
174	95	50	100	97.2	45307	PASS
175	174	5	9	7.4	3343	PASS
176	174	95	101	95.0	43040	PASS
177	176	5	9	6.5	2816	PASS