


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv030920\
 Data File : VY001903.D
 Acq On : 09 Mar 2020 09:29
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
 Title : SW846 8260
 Last Update : Thu Mar 05 04:05:33 2020



AutoFind: Scans 1763, 1764, 1765; Background Corrected with Scan 1755

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	26466	PASS
75	95	30	60	51.4	72258	PASS
95	95	100	100	100.0	140488	PASS
96	95	5	9	6.7	9412	PASS
173	174	0.00	2	0.9	875	PASS
174	95	50	100	69.0	96928	PASS
175	174	5	9	7.4	7191	PASS
176	174	95	101	97.1	94122	PASS
177	176	5	9	6.9	6466	PASS