

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031023\
 Data File : VY012918.D
 Acq On : 10 Mar 2023 16:58
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 11 05:49:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.493	0.503	-2.0	87	0.00
3 P	Chloromethane	0.589	0.572	2.9	86	0.00
4 C	Vinyl Chloride	0.540	0.584	-8.1#	90	0.00
5 T	Bromomethane	0.328	0.345	-5.2	89	0.00
6 T	Chloroethane	0.324	0.352	-8.6	89	0.00
7 T	Trichlorofluoromethane	0.897	0.994	-10.8	93	0.00
8 T	Diethyl Ether	0.314	0.327	-4.1	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.604	-11.9	93	0.00
10 T	Methyl Iodide	0.703	0.781	-11.1	86	0.00
11 T	Tert butyl alcohol	0.070	0.052	25.7#	75	0.00
12 CM	1,1-Dichloroethene	0.515	0.565	-9.7#	90	0.00
13 T	Acrolein	0.087	0.056	35.6#	53	0.00
14 T	Allyl chloride	0.951	1.036	-8.9	88	0.00
15 T	Acrylonitrile	0.166	0.170	-2.4	83	0.00
16 T	Acetone	0.195	0.197	-1.0	76	0.00
17 T	Carbon Disulfide	1.753	1.774	-1.2	84	0.00
18 T	Methyl Acetate	0.580	0.553	4.7	77	0.00
19 T	Methyl tert-butyl Ether	1.330	1.491	-12.1	89	0.00
20 T	Methylene Chloride	0.714	0.682	4.5	91	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.642	-14.6	94	0.00
22 T	Diisopropyl ether	1.727	2.053	-18.9	95	0.00
23 T	Vinyl Acetate	1.073	1.227	-14.4	89	0.00
24 P	1,1-Dichloroethane	0.984	1.160	-17.9	96	0.00
25 T	2-Butanone	0.241	0.246	-2.1	79	0.00
26 T	2,2-Dichloropropane	0.874	1.031	-18.0	98	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.702	-17.0	94	0.00
28 T	Bromochloromethane	0.360	0.397	-10.3	88	0.00
29 T	Tetrahydrofuran	0.146	0.147	-0.7	79	0.00
30 C	Chloroform	0.972	1.178	-21.2#	98	0.00
31 T	Cyclohexane	0.951	0.974	-2.4	86	0.00
32 T	1,1,1-Trichloroethane	0.879	1.001	-13.9	93	0.00
33 S	1,2-Dichloroethane-d4	0.496	0.466	6.0	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.302	0.313	-3.6	83	0.00
36 T	1,1-Dichloropropene	0.506	0.578	-14.2	91	0.00
37 T	Ethyl Acetate	0.319	0.338	-6.0	87	0.00
38 T	Carbon Tetrachloride	0.522	0.620	-18.8	94	0.00
39 T	Methylcyclohexane	0.601	0.682	-13.5	87	0.00
40 TM	Benzene	1.430	1.645	-15.0	91	0.00
41 T	Methacrylonitrile	0.172	0.207	-20.3	94	0.01
42 TM	1,2-Dichloroethane	0.430	0.487	-13.3	89	0.00
43 T	Isopropyl Acetate	0.569	0.585	-2.8	80	0.00
44 TM	Trichloroethene	0.377	0.434	-15.1	90	0.00
45 C	1,2-Dichloropropane	0.361	0.414	-14.7#	92	0.00
46 T	Dibromomethane	0.206	0.230	-11.7	87	0.00
47 T	Bromodichloromethane	0.485	0.560	-15.5	91	0.00
48 T	Methyl methacrylate	0.262	0.268	-2.3	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031023\
 Data File : VY012918.D
 Acq On : 10 Mar 2023 16:58
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 11 05:49:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	80	0.00
50 S	Toluene-d8	1.103	1.084	1.7	79	0.00
51 T	4-Methyl-2-Pentanone	0.308	0.315	-2.3	77	0.00
52 CM	Toluene	0.863	1.025	-18.8#	93	0.00
53 T	t-1,3-Dichloropropene	0.509	0.586	-15.1	89	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.663	-15.1	89	0.00
55 T	1,1,2-Trichloroethane	0.287	0.319	-11.1	88	0.00
56 T	Ethyl methacrylate	0.382	0.423	-10.7	82	0.00
57 T	1,3-Dichloropropane	0.489	0.548	-12.1	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.207	-3.0	77	0.00
59 T	2-Hexanone	0.224	0.230	-2.7	76	0.00
60 T	Dibromochloromethane	0.349	0.405	-16.0	91	0.00
61 T	1,2-Dibromoethane	0.276	0.304	-10.1	86	0.00
62 S	4-Bromofluorobenzene	0.372	0.378	-1.6	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.364	0.421	-15.7	92	0.00
65 PM	Chlorobenzene	1.049	1.202	-14.6	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.466	-20.1	95	0.00
67 C	Ethyl Benzene	1.845	2.201	-19.3#	92	0.00
68 T	m/p-Xylenes	0.710	0.849	-19.6	92	0.00
69 T	o-Xylene	0.662	0.790	-19.3	93	0.00
70 T	Styrene	1.105	1.343	-21.5	92	0.00
71 P	Bromoform	0.253	0.281	-11.1	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.544	4.162	-17.4	94	0.00
74 T	N-amyl acetate	1.115	1.117	-0.2	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.797	0.830	-4.1	86	0.00
76 T	1,2,3-Trichloropropane	0.597	0.521	12.7	72	0.00
77 T	Bromobenzene	0.861	0.967	-12.3	92	0.00
78 T	n-propylbenzene	4.369	5.097	-16.7	93	0.00
79 T	2-Chlorotoluene	2.428	2.793	-15.0	93	0.00
80 T	1,3,5-Trimethylbenzene	2.924	3.397	-16.2	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.288	0.288	0.0	80	0.00
82 T	4-Chlorotoluene	2.513	2.861	-13.8	92	0.00
83 T	tert-Butylbenzene	2.524	3.061	-21.3	96	0.00
84 T	1,2,4-Trimethylbenzene	2.860	3.364	-17.6	94	0.00
85 T	sec-Butylbenzene	3.851	4.530	-17.6	95	0.00
86 T	p-Isopropyltoluene	3.147	3.706	-17.8	94	0.00
87 T	1,3-Dichlorobenzene	1.660	1.895	-14.2	94	0.00
88 T	1,4-Dichlorobenzene	1.694	1.888	-11.5	92	0.00
89 T	n-Butylbenzene	3.013	3.519	-16.8	94	0.00
90 T	Hexachloroethane	0.655	0.751	-14.7	96	0.00
91 T	1,2-Dichlorobenzene	1.501	1.678	-11.8	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.129	0.129	0.0	80	0.00
93 T	1,2,4-Trichlorobenzene	0.924	1.012	-9.5	88	0.00
94 T	Hexachlorobutadiene	0.564	0.630	-11.7	95	0.00
95 T	Naphthalene	1.852	1.942	-4.9	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031023\
Data File : VY012918.D
Acq On : 10 Mar 2023 16:58
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 11 05:49:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.822	0.883	-7.4	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6