

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031023\
 Data File : VY012914.D
 Acq On : 10 Mar 2023 12:33
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Mar 11 05:48:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	87785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	145520	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	136960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	64820	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	54390	62.403	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.800%
35) Dibromofluoromethane	7.716	113	48955	55.744	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.480%
50) Toluene-d8	10.179	98	176370	54.934	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.860%
62) 4-Bromofluorobenzene	12.483	95	59334	54.754	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.500%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031023\
Data File : VY012914.D
Acq On : 10 Mar 2023 12:33
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Mar 11 05:48:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY012914.D
Acq: 10 Mar 2023 12:33




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 87785
Ion Ratio Lower Upper
168 100
99 50.7 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 62.403 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.000 min
Lab File: VY012914.D
Acq: 10 Mar 2023 12:33

Tgt Ion: 65 Resp: 54390
Ion Ratio Lower Upper
65 100
67 53.2 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY012914.D

Acq: 10 Mar 2023 12:33

Instrument :

MSVOA_Y

ClientSampleId :

VIBLK





Tgt Ion:114 Resp: 145520

Ion Ratio Lower Upper

114 100

63 19.7 0.0 37.4

88 15.0 0.0 28.6



#35

Dibromofluoromethane

Concen: 55.744 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY012914.D

Acq: 10 Mar 2023 12:33




Tgt Ion:113 Resp: 48955


Ion Ratio Lower Upper

113 100

111 102.7 82.2 123.2

192 21.5 16.8 25.2









#50
Toluene-d8
Concen: 54.934 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY012914.D
Acq: 10 Mar 2023 12:33




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion: 98 Resp: 176370
Ion Ratio Lower Upper
98 100
100 63.0 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 54.754 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY012914.D
Acq: 10 Mar 2023 12:33

Tgt Ion: 95 Resp: 59334
Ion Ratio Lower Upper
95 100
174 88.3 0.0 176.6
176 85.0 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1158
Delta R.T. -0.000 min
Lab File: VY012914.D
Acq: 10 Mar 2023 12:33

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Tgt Ion:117 Resp: 136960
Ion Ratio Lower Upper
117 100
82 55.8 45.4 68.0
119 32.4 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY012914.D
Acq: 10 Mar 2023 12:33

Tgt Ion:152 Resp: 64820
Ion Ratio Lower Upper
152 100
115 57.6 27.1 81.3
150 158.4 0.0 347.8

