

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031120\
 Data File : VY001934.D
 Acq On : 11 Mar 2020 11:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 12 09:58:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 04:05:33 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.462	0.427	7.6	85	0.00
3 P	Chloromethane	0.622	0.612	1.6	89	0.00
4 C	Vinyl Chloride	0.633	0.638	-0.8#	88	0.00
5 T	Bromomethane	0.420	0.423	-0.7	91	0.00
6 T	Chloroethane	0.398	0.415	-4.3	92	0.00
7 T	Trichlorofluoromethane	0.858	0.879	-2.4	92	0.00
8 T	Diethyl Ether	0.338	0.358	-5.9	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.552	0.576	-4.3	94	0.00
10 T	Methyl Iodide	0.638	0.670	-5.0	85	-0.01
11 T	Tert butyl alcohol	0.065	0.057	12.3	88	-0.02
12 CM	1,1-Dichloroethene	0.556	0.582	-4.7#	92	-0.01
13 T	Acrolein	0.075	0.076	-1.3	99	0.00
14 T	Allyl chloride	0.986	1.055	-7.0	93	-0.01
15 T	Acrylonitrile	0.181	0.184	-1.7	89	0.00
16 T	Acetone	0.155	0.158	-1.9	88	0.00
17 T	Carbon Disulfide	1.869	1.851	1.0	87	-0.01
18 T	Methyl Acetate	0.399	0.392	1.8	86	-0.01
19 T	Methyl tert-butyl Ether	1.506	1.579	-4.8	91	-0.01
20 T	Methylene Chloride	0.666	0.670	-0.6	94	-0.01
21 T	trans-1,2-Dichloroethene	0.631	0.661	-4.8	92	0.00
22 T	Diisopropyl ether	2.020	2.177	-7.8	94	0.00
23 T	Vinyl Acetate	1.324	1.410	-6.5	92	-0.01
24 P	1,1-Dichloroethane	1.094	1.154	-5.5	93	-0.01
25 T	2-Butanone	0.236	0.242	-2.5	88	0.00
26 T	2,2-Dichloropropane	0.898	0.956	-6.5	95	-0.01
27 T	cis-1,2-Dichloroethene	0.686	0.728	-6.1	92	-0.01
28 T	Bromochloromethane	0.493	0.515	-4.5	91	0.00
29 T	Tetrahydrofuran	0.157	0.157	0.0	88	-0.01
30 C	Chloroform	1.028	1.093	-6.3#	93	-0.01
31 T	Cyclohexane	1.173	1.169	0.3	92	0.00
32 T	1,1,1-Trichloroethane	0.866	0.923	-6.6	92	0.00
33 S	1,2-Dichloroethane-d4	0.534	0.530	0.7	92	-0.01
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.273	0.284	-4.0	94	-0.01
36 T	1,1-Dichloropropene	0.481	0.509	-5.8	91	0.00
37 T	Ethyl Acetate	0.280	0.300	-7.1	91	0.00
38 T	Carbon Tetrachloride	0.411	0.441	-7.3	93	0.00
39 T	Methylcyclohexane	0.638	0.668	-4.7	92	-0.01
40 TM	Benzene	1.426	1.527	-7.1	93	-0.01
41 T	Methacrylonitrile	0.126	0.121	4.0	86	-0.02
42 TM	1,2-Dichloroethane	0.367	0.389	-6.0	91	0.00
43 T	Isopropyl Acetate	0.529	0.557	-5.3	90	0.00
44 TM	Trichloroethene	0.347	0.369	-6.3	93	-0.01
45 C	1,2-Dichloropropane	0.359	0.387	-7.8#	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031120\
 Data File : VY001934.D
 Acq On : 11 Mar 2020 11:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 12 09:58:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 04:05:33 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.185	0.195	-5.4	92	-0.01
47 T	Bromodichloromethane	0.430	0.474	-10.2	94	-0.01
48 T	Methyl methacrylate	0.235	0.248	-5.5	86	0.00
49 T	1,4-Dioxane	0.003	0.002	33.3#	80	-0.02
50 S	Toluene-d8	1.151	1.182	-2.7	93	0.00
51 T	4-Methyl-2-Pentanone	0.273	0.287	-5.1	90	0.00
52 CM	Toluene	0.866	0.931	-7.5#	93	0.00
53 T	t-1,3-Dichloropropene	0.468	0.521	-11.3	93	0.00
54 T	cis-1,3-Dichloropropene	0.557	0.613	-10.1	93	0.00
55 T	1,1,2-Trichloroethane	0.270	0.285	-5.6	91	0.00
56 T	Ethyl methacrylate	0.396	0.423	-6.8	89	0.00
57 T	1,3-Dichloropropane	0.482	0.513	-6.4	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.157	0.133	15.3	85	0.00
59 T	2-Hexanone	0.195	0.201	-3.1	87	0.00
60 T	Dibromochloromethane	0.285	0.315	-10.5	93	0.00
61 T	1,2-Dibromoethane	0.256	0.270	-5.5	92	0.00
62 S	4-Bromofluorobenzene	0.400	0.413	-3.2	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.314	0.325	-3.5	94	0.00
65 PM	Chlorobenzene	0.996	1.040	-4.4	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.327	0.352	-7.6	94	0.00
67 C	Ethyl Benzene	1.858	1.968	-5.9#	93	0.00
68 T	m/p-Xylenes	0.687	0.737	-7.3	94	0.00
69 T	o-Xylene	0.647	0.687	-6.2	93	0.00
70 T	Styrene	1.109	1.194	-7.7	92	0.00
71 P	Bromoform	0.177	0.191	-7.9	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.817	4.049	-6.1	92	0.00
74 T	N-amyl acetate	1.173	1.217	-3.8	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.823	0.854	-3.8	90	0.00
76 T	1,2,3-Trichloropropane	0.593	0.615	-3.7	97	0.00
77 T	Bromobenzene	0.809	0.870	-7.5	95	0.00
78 T	n-propylbenzene	4.709	5.069	-7.6	94	0.00
79 T	2-Chlorotoluene	2.602	2.777	-6.7	93	0.00
80 T	1,3,5-Trimethylbenzene	3.142	3.377	-7.5	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.294	0.315	-7.1	90	0.00
82 T	4-Chlorotoluene	2.709	2.894	-6.8	93	0.00
83 T	tert-Butylbenzene	2.610	2.828	-8.4	95	0.00
84 T	1,2,4-Trimethylbenzene	3.118	3.356	-7.6	93	0.00
85 T	sec-Butylbenzene	3.847	4.218	-9.6	95	0.00
86 T	p-Isopropyltoluene	3.318	3.613	-8.9	94	0.00
87 T	1,3-Dichlorobenzene	1.585	1.688	-6.5	94	0.00
88 T	1,4-Dichlorobenzene	1.608	1.700	-5.7	93	0.00
89 T	n-Butylbenzene	3.443	3.747	-8.8	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031120\
Data File : VY001934.D
Acq On : 11 Mar 2020 11:25
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 12 09:58:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 05 04:05:33 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.695	-11.0	96	0.00
91 T	1,2-Dichlorobenzene	1.440	1.545	-7.3	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.130	-1.6	91	0.00
93 T	1,2,4-Trichlorobenzene	0.934	0.974	-4.3	91	0.00
94 T	Hexachlorobutadiene	0.455	0.486	-6.8	96	0.00
95 T	Naphthalene	2.215	2.267	-2.3	90	0.00
96 T	1,2,3-Trichlorobenzene	0.822	0.859	-4.5	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6