


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031120\
 Data File : VY001933.D
 Acq On : 11 Mar 2020 10:51
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
 Title : SW846 8260
 Last Update : Thu Mar 05 04:05:33 2020



AutoFind: Scans 1762, 1763, 1764; Background Corrected with Scan 1754

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	29042	PASS
75	95	30	60	51.5	76688	PASS
95	95	100	100	100.0	148802	PASS
96	95	5	9	6.8	10067	PASS
173	174	0.00	2	0.9	948	PASS
174	95	50	100	68.2	101557	PASS
175	174	5	9	7.6	7708	PASS
176	174	95	101	96.6	98077	PASS
177	176	5	9	7.0	6878	PASS