



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031224\
Data File : VY017636.D
Acq On : 12 Mar 2024 08:27
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Title : SW846 8260
Last Update : Sat Feb 24 04:41:20 2024



AutoFind: Scans 1750, 1751, 1752; Background Corrected with Scan 1742

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	7759	PASS
75	95	30	60	49.9	20336	PASS
95	95	100	100	100.0	40749	PASS
96	95	5	9	6.8	2776	PASS
173	174	0.00	2	1.3	501	PASS
174	95	50	100	93.3	38029	PASS
175	174	5	9	7.2	2745	PASS
176	174	95	101	96.7	36792	PASS
177	176	5	9	6.5	2382	PASS