

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031323\
 Data File : VY012924.D
 Acq On : 13 Mar 2023 13:22
 Operator : KP/MD
 Sample : 01877-02
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-SOIL

Quant Time: Mar 14 01:26:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	82959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	139108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	126389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	56377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	50120	60.849	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.700%
35) Dibromofluoromethane	7.716	113	46680	55.604	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.200%
50) Toluene-d8	10.179	98	165566	53.946	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.900%
62) 4-Bromofluorobenzene	12.483	95	49455	47.741	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.480%
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	4775	14.770	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031323\
Data File : VY012924.D
Acq On : 13 Mar 2023 13:22
Operator : KP/MD
Sample : 01877-02
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-SOIL

Quant Time: Mar 14 01:26:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

