

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031323\
 Data File : VY012925.D
 Acq On : 13 Mar 2023 14:25
 Operator : KP/MD
 Sample : 01863-03
 Misc : 5.13g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 030923-1-VOC

Quant Time: Mar 14 01:26:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	84191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	142848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	127414	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	56775	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	51770	61.933	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.860%
35) Dibromofluoromethane	7.716	113	48330	56.062	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.120%
50) Toluene-d8	10.185	98	170358	54.054	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.100%
62) 4-Bromofluorobenzene	12.483	95	50130	47.126	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031323\
Data File : VY012925.D
Acq On : 13 Mar 2023 14:25
Operator : KP/MD
Sample : 01863-03
Misc : 5.13g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
030923-1-VOC

Quant Time: Mar 14 01:26:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

