

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031323\
 Data File : VY012921.D
 Acq On : 13 Mar 2023 11:47
 Operator : KP/MD
 Sample : VY0313SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0313SBL01

Quant Time: Mar 14 01:24:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	87146	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	146180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	133242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	59828	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	51736	59.793	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.580%
35) Dibromofluoromethane	7.716	113	48290	54.739	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.480%
50) Toluene-d8	10.179	98	171807	53.271	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.540%
62) 4-Bromofluorobenzene	12.483	95	52679	48.393	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.780%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031323\
Data File : VY012921.D
Acq On : 13 Mar 2023 11:47
Operator : KP/MD
Sample : VY0313SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0313SBL01

Quant Time: Mar 14 01:24:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY012921.D
Acq: 13 Mar 2023 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VY0313SBL01


Tgt Ion:168 Resp: 87146
Ion Ratio Lower Upper
168 100
99 53.5 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 59.793 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.000 min
Lab File: VY012921.D
Acq: 13 Mar 2023 11:47

Tgt Ion: 65 Resp: 51736
Ion Ratio Lower Upper
65 100
67 52.7 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012921.D

Acq: 13 Mar 2023 11:47

Instrument :

MSVOA_Y

ClientSampleId :

VY0313SBL01





Tgt Ion:114 Resp: 146180

Ion Ratio Lower Upper

114 100

63 18.4 0.0 37.4

88 14.2 0.0 28.6



#35

Dibromofluoromethane

Concen: 54.739 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY012921.D

Acq: 13 Mar 2023 11:47




Tgt Ion:113 Resp: 48290


Ion Ratio Lower Upper

113 100

111 102.5 82.2 123.2

192 21.3 16.8 25.2









#50
Toluene-d8
Concen: 53.271 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY012921.D
Acq: 13 Mar 2023 11:47




Instrument :
MSVOA_Y
ClientSampleId :
VY0313SBL01


Tgt Ion: 98 Resp: 171807
Ion Ratio Lower Upper
98 100
100 63.9 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 48.393 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY012921.D
Acq: 13 Mar 2023 11:47

Tgt Ion: 95 Resp: 52679
Ion Ratio Lower Upper
95 100
174 93.0 0.0 176.6
176 88.6 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY012921.D
Acq: 13 Mar 2023 11:47

Instrument :
MSVOA_Y
ClientSampleId :
VY0313SBL01

Tgt Ion:117 Resp: 133242
Ion Ratio Lower Upper
117 100
82 53.7 45.4 68.0
119 32.2 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY012921.D
Acq: 13 Mar 2023 11:47

Tgt Ion:152 Resp: 59828
Ion Ratio Lower Upper
152 100
115 54.7 27.1 81.3
150 155.8 0.0 347.8

