

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
 Data File : VY012940.D
 Acq On : 14 Mar 2023 14:13
 Operator : KP/MD
 Sample : 01905-07
 Misc : 5.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MAPLE-2

Quant Time: Mar 15 01:42:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	80423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	134329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	125544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	57326	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	54553	68.320	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.640%
35) Dibromofluoromethane	7.716	113	47019	58.000	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.000%
50) Toluene-d8	10.179	98	164745	55.588	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.180%
62) 4-Bromofluorobenzene	12.483	95	50052	50.037	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
Data File : VY012940.D
Acq On : 14 Mar 2023 14:13
Operator : KP/MD
Sample : 01905-07
Misc : 5.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MAPLE-2

Quant Time: Mar 15 01:42:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

