

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
 Data File : VY012947.D
 Acq On : 14 Mar 2023 17:20
 Operator : KP/MD
 Sample : 01908-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HILLBOR-DRUMS-0314

Quant Time: Mar 15 01:45:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	37013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	59387	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	53054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	21961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	22410	60.981	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.960%	
35) Dibromofluoromethane	7.716	113	21577	60.204	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 120.400%	
50) Toluene-d8	10.179	98	70551	53.846	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.700%	
62) 4-Bromofluorobenzene	12.483	95	19616	44.356	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 88.720%	
Target Compounds						
30) Chloroform	7.508	83	1741	2.419	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
Data File : VY012947.D
Acq On : 14 Mar 2023 17:20
Operator : KP/MD
Sample : 01908-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HILLBOR-DRUMS-0314

Quant Time: Mar 15 01:45:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

